



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not  
Complete this report whenever the instrument is serviced or repaired and whenever it is pla  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**RECEIVED**

By Carol Day at 2:56 pm, Sep 02, 2015

DATAMASTER SN 201209	NAME OF AGENCY Perryville PD	DATE OF INSPECTION 09/02/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 120 N. Jackson St., Perryville, MO		TIME OF INSPECTION 1:43 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 09/02/2015 1343
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Repco Manufacturing LOT # 15001 EXP. DATE 05/20/2017	
<input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN sd2778 EXP. DATE 05/05/2016	
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 .103	TEST 2 .102	TEST 3 .103
-------------	-------------	-------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS  
(USE OTHER SIDE IF NECESSARY).

Instrument is working within DHSS Guidelines

**INSPECTING OFFICER**

SIGNATURE <i>Ryan L. Worthington</i>	PRINT FULL NAME Ryan L. Worthington
TYPE II PERMID NUMBER/EXPIRATION DATE 250120 05/14/2017	TELEPHONE NUMBER (573) 547-4546

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 15001**

**EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President

RepCo Marketing, Inc.

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
BAC DATAMASTER SERIAL NUMBER 801209  
09/02/15

ARREST TIME: 13:30

SUBJECT NAME:

DOE/JON/B

DOB: 01/01/01 SEX: M

STATE/D.L.: MO/1234567890

ARRESTING OFFICER:

WORTHINGTON/R/L

OFFICER I.D.: 105

TESTING OFFICER:

SAME

OFFICER I.D.: SAME

PERMIT NUMBER: 250120

EXPIRATION DATE: 05/14/17


MISCELLANEOUS DATA:

RFI TEST

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature



# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
BAC DATA MASTER SERIAL NUMBER 201209  
09/02/15

TESTING OFFICER:  
BORTHINGTON/RYAN/L  
OFFICER J.D.: 105  
PERMIT NUMBER: 250120  
EXPIRATION DATE: 05/14/17  
MISCELLANEOUS DATA:

## --- SUPERVISOR MODE ---

BLANK TEST	.000	13:50
INTERNAL STANDARD	VERIFIED	13:50
EXTERNAL STANDARD	.103	13:50
BLANK TEST	.000	13:51
EXTERNAL STANDARD	.102	13:52
BLANK TEST	.000	13:52
EXTERNAL STANDARD	.103	13:53
BLANK TEST	.000	13:53

N = 3  
STDEV. = .1  
AVG. = .1026

Operator Signature *R. J. Borthington*  
Printed on recycled paper with agri-based inks

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
BAC DATA MASTER SERIAL NUMBER 201209  
09/02/15  
13:43

## --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HERTERS  
SAMPLE CHAMBER: 48C  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
WARRANTY STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~

Operator Signature *R. J. Borthington*  
Printed on recycled paper with agri-based inks



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**RYAN L WORTHINGTON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX DMT**


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/14/2015

NUMBER 250120

EXPIRES 5/14/2017

  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY



DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator WORTHINGTON, RYAN  
Permit No 250120  
Date Issued 5/14/2015 Date Expires 5/14/2017